

EC DECLARATION OF CONFORMITY

Document number: RAE/NGA2000FRC-E6

Date: June 2003

We,

Emerson Process Management Manufacturing GmbH & Co. OHG

located at

Industriestraße 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

NGA2000 FRC

to which this declaration relates is in conformity with the provisions of:

89/336/EEC EMC Directive (changed by directive 91/263/EEC 92/31/EEC and 93/68/EEC)
with the application of the harmonized standards:

EN 61326-1:1997 Electrical equipment for measurement, control and laboratory use - EMC requirements
+ A1:1998 Part 1: General requirements

97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3, paragraph 3 of the above mentioned directive and therefore CE marking does not refer to this directive.

NGA 2000 FRC does include any ½ 19" NGA 2000 FRC analyzer (both rack-mountable or table-top version).

Hasselroth, 24.06.2003



.....

(Signature)

Ian Macleod

(Name)

VP Sales & Marketing EMA

(Function name)